



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

EPPENDORF SAN DIEGO SERVICE CENTER
7270 Engineer Road Suite A
San Diego, CA 92111
Marcela Sanclemente Phone: 516 515 2669

CALIBRATION

Valid To: December 31, 2020

Certificate Number: 3263.02

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Fluid Quantities

Parameter/Equipment	Range	CMC ² (±)	Comments
Pipettes – Repeaters, Bottle Tops, Top Burets	(0.1 to 10) µL (10 to 100) µL (100 to 1000) µL (1000 to 10 000) µL (10 000 to 25 000) µL (25 000 to 50 000) µL (50 000 to 100 000) µL	0.032 µL 0.19 µL 1.3 µL 13 µL 100 µL 120 µL 150 µL	Gravimetric method referenced to mass balances and ASTM Class 1 mass standards

¹ This laboratory offers commercial calibration service.

² Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMC's represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.



Accredited Laboratory

A2LA has accredited

EPPENDORF SAN DIEGO SERVICE CENTER

San Diego, CA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General requirements for the competence of testing and calibration laboratories*. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 6th day of November 2018.

A handwritten signature in black ink, written over a horizontal line.

President and CEO
For the Accreditation Council
Certificate Number 3263.02
Valid to December 31, 2020

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.